

**E46C02 52.246-9003 MEASURING AND TEST EQUIPMENT (JUN 1998)**

**DLAD**

Old code: E17

Notwithstanding any other clause to the contrary, and/or in addition thereto, the contractor shall ensure that the gauges and other measuring and testing equipment, used in determining whether the supplies presented to the Government for acceptance under this contract fully conform to specified technical requirements, are calibrated in accordance with ISO 10012-1 or ANSI/NCLS Z540-1.

(End of Clause)